# MEMBRAPOR SPECIFICATION SHEET



# H2O2/MB-100

## Hydrogen Peroxide Gas Sensor in Mini Housing

#### **MEASUREMENT**

Operation Principle	3-Electrode Electrochemical	
Nominal Range	0 – 100 ppm	
Maximum Overload	200 ppm	
Inboard Filter	_	
Output Signal	900 ± 200 nA/ppm	
Resolution (Electronics dependent)	< 0.3 ppm	
T90 Response Time	< 60 sec	
Typical Baseline Range (pure air, 20°C)	0 ppm to 2 ppm	
Maximum Zero Shift (+20°C to +40°C)	N.D.	
Repeatability	< 2 % of signal	
Output Linearity	Linear	
Gain	_	

#### **ELECTRICAL**

Rec. Load Resistor	10 Ohm
Bias (V_Sens-V_Ref)	+300 mV
Conformity to RoHS directive	RoHS Compliance

#### **ENVIRONMENTAL**


Relative Humidity Range	15 % to 90 % R.H. non- condensing
Temperature Range	-20 °C to 50 °C
Pressure Range	Atmospheric ± 10%
Pressure Coefficient	N.D.
Humidity Effect	none

#### **LIFETIME**

Expected Operation Life	2 years in air
Expected Long Term Output Drift in air	N.D.
Filter Life	_
Storage Life	6 months in container
Rec. Storage Temperature	5 °C – 20 °C
Warranty Period	12 months from date of dispatch

Performance data conditions: 20 °C, 50% RH, 1013 mbar

#### Miniature-Size Outline Dimensions



± 0.10 mm

#### **MECHANICAL**

Weight	5.5 g
Position Sensitivity	None

## **APPLICATIONS**

Discontinuous Measurement Sterilization Processes

#### **CROSS-SENSITIVITY DATA**

The table below does not claim to be complete. Interfering gases should not be used for calibration.

Interfering Gas	Conc.	Reading
	ppm	ppm
SO <sub>2</sub>	100	100 ± 60

REV.: 01/2018 Page 1 of 1

Phone: +41 43 311 72 00

Fax: +41 43 311 72 01

Email: info@membrapor.ch

MEMBRAPOR AG

Birkenweg 2

CH-8304 Wallisellen

www.membrapor.ch

Switzerland

The data contained in this document is for guidance only. Membrapor AG accepts no liability for any consequential losses, injury or damage resulting from the use of this document or from any omissions or errors herein. Customers should test under their own conditions, to ensure that the sensors are suitable for their own requirements.