

Technical specifications:

G460



Measuring principle	Electrochemical (EC): for toxic gases and oxygen Photoionisation (EC): for toxic combustible vapours (in ppm range) Catalytic combustion (CC): for flammable gases and vapors (up to 100%LEL) Infrared (IR): for flammable gases and vapors and carbon dioxide												
Measuring ranges	sensor dependent												
Response time	sensor dependent												
Expected average life of the sensor	sensor dependent												
Measuring gas supply	Diffusion with flow velocity of 0...6 m/s or pump by means of attachable electrical sampling pump G400-MP2												
Display	illuminated LCD full graphics display, automatic size setting for optimum reading, displays the battery capacity, gas concentration as current value and peak value												
Alerting	depending on the gas type 3 or 2 momentary value and 2 exposure level alarms, battery alarm with visual and acoustical signaling as well as display on the screen, color of the display depending on the alarm state (orange/red). Horn: 103 dB(A) (can be reduced to 90 dB(A))												
Zero point and sensitivity adjustment	manual or automatic with an adjustment program, if necessary, test gas supply via the "SMART CAP" or the „SMART CHARGER CAP" with 0.5...0.6slpm												
Power supply	1. NiMH battery module A21 (colour: black), 2100mAh rechargeable 2. NiMH battery module F25 (colour: black), 2500mAh rechargeable Im=600mA (max. charging current) Um=6V DC (max. voltage) or 3. Alkaline battery module (colour grey), non-rechargeable with 2x Mignon 1,5V Type: DURACELL PROCELL MN1500 LR6 AA or INDUSTRIAL BY DURACELL ID1500 AA (LR6)												
Operating time (*1)	<table border="1"> <tr> <td>NiMH-II A21:</td> <td>approx. 9h (EC+CC); approx. 7h (EC+IR+CC); approx. 8h (EC+PID+CC); approx. 6h (EC+PID+IR+CC);</td> <td>approx. 17h (EC+CC_{CH4}); approx. 11h (EC+IR+CC_{CH4}); approx. 12h (EC+PID+CC_{CH4}); approx. 8h (EC+PID+IR+CC_{CH4});</td> <td>approx. 120h (EC) approx. 25h (EC+IR) approx. 33h (EC+PID) approx. 14h (EC+PID+IR)</td> </tr> <tr> <td>NiMH F25:</td> <td>approx. 11h (EC+CC); approx. 8h (EC+IR+CC); approx. 9h (EC+PID+CC); approx. 7h (EC+PID+IR+CC);</td> <td>approx. 20h (EC+CC_{CH4}); approx. 12h (EC+IR+CC_{CH4}); approx. 14h (EC+PID+CC_{CH4}); approx. 10h (EC+PID+IR+CC_{CH4});</td> <td>approx. 130h (EC) approx. 30h (EC+IR) approx. 40h (EC+PID) approx. 17h (EC+PID+IR)</td> </tr> <tr> <td>Alkaline:</td> <td>approx. 8h (EC+CC); approx. 5h (EC+IR+CC); approx. 6h (EC+PID+CC);</td> <td>approx. 14h (EC+CC_{CH4}); approx. 6h (EC+IR+CC_{CH4}); approx. 9h (EC+PID+CC_{CH4});</td> <td>approx. 170h (EC) approx. 28h (EC+IR) approx. 40h (EC+PID)</td> </tr> </table>	NiMH-II A21:	approx. 9h (EC+CC); approx. 7h (EC+IR+CC); approx. 8h (EC+PID+CC); approx. 6h (EC+PID+IR+CC);	approx. 17h (EC+CC _{CH4}); approx. 11h (EC+IR+CC _{CH4}); approx. 12h (EC+PID+CC _{CH4}); approx. 8h (EC+PID+IR+CC _{CH4});	approx. 120h (EC) approx. 25h (EC+IR) approx. 33h (EC+PID) approx. 14h (EC+PID+IR)	NiMH F25:	approx. 11h (EC+CC); approx. 8h (EC+IR+CC); approx. 9h (EC+PID+CC); approx. 7h (EC+PID+IR+CC);	approx. 20h (EC+CC _{CH4}); approx. 12h (EC+IR+CC _{CH4}); approx. 14h (EC+PID+CC _{CH4}); approx. 10h (EC+PID+IR+CC _{CH4});	approx. 130h (EC) approx. 30h (EC+IR) approx. 40h (EC+PID) approx. 17h (EC+PID+IR)	Alkaline:	approx. 8h (EC+CC); approx. 5h (EC+IR+CC); approx. 6h (EC+PID+CC);	approx. 14h (EC+CC _{CH4}); approx. 6h (EC+IR+CC _{CH4}); approx. 9h (EC+PID+CC _{CH4});	approx. 170h (EC) approx. 28h (EC+IR) approx. 40h (EC+PID)
NiMH-II A21:	approx. 9h (EC+CC); approx. 7h (EC+IR+CC); approx. 8h (EC+PID+CC); approx. 6h (EC+PID+IR+CC);	approx. 17h (EC+CC _{CH4}); approx. 11h (EC+IR+CC _{CH4}); approx. 12h (EC+PID+CC _{CH4}); approx. 8h (EC+PID+IR+CC _{CH4});	approx. 120h (EC) approx. 25h (EC+IR) approx. 33h (EC+PID) approx. 14h (EC+PID+IR)										
NiMH F25:	approx. 11h (EC+CC); approx. 8h (EC+IR+CC); approx. 9h (EC+PID+CC); approx. 7h (EC+PID+IR+CC);	approx. 20h (EC+CC _{CH4}); approx. 12h (EC+IR+CC _{CH4}); approx. 14h (EC+PID+CC _{CH4}); approx. 10h (EC+PID+IR+CC _{CH4});	approx. 130h (EC) approx. 30h (EC+IR) approx. 40h (EC+PID) approx. 17h (EC+PID+IR)										
Alkaline:	approx. 8h (EC+CC); approx. 5h (EC+IR+CC); approx. 6h (EC+PID+CC);	approx. 14h (EC+CC _{CH4}); approx. 6h (EC+IR+CC _{CH4}); approx. 9h (EC+PID+CC _{CH4});	approx. 170h (EC) approx. 28h (EC+IR) approx. 40h (EC+PID)										
Climatic conditions	for operation: -20...+50°C 5...95%r.h. 70...130kPa for storage: -25...+60°C 5...95%r.h. 70...130kPa (recommended 0...+30°C)												
Housing	Material: rubberized plastic Dimensions: 75 x 110 x 55 mm (W x H x D) Weight: up to 350 g (depending on sensor configuration) Protection class: IP67												

Technical specifications:

G460



Approvals / Tests

Markings and ignition protection types:	⊕ II 2G	Ex ia d IIC T4 Gb	-20°C ≤ Ta ≤ +55°C	for NiMH-II (black)
		Ex ia d IIC T3 Gb	-20°C ≤ Ta ≤ +55°C	for NiMH (black)
		Ex ia d IIC T4/T3 Gb	-20°C ≤ Ta ≤ +45°C/+55°C	for Alkaline (grey)
	⊕ I M2	Ex ia d I Mb	-20°C ≤ Ta ≤ +55°C	for NiMH, NiMH-II (black)
EU Type Examination Certificate:	BVS 06 ATEX E 017 X	(for measuring function and electronic Ex-protection)		
EU Type Examination Certificate:	PFG 09 G 001	(for measuring function)		
Electromagnetic compability:	DIN EN 50270:2006	Interference emission: Type class I Interference immunity: Type class II		

to (*1): The service life is indicated for new battery modules at operating temperatures of +20°C. It will be reduced by pressing buttons (display lighting & lamp) and by gas alarms. It is reduced with the age of the battery module, with the number of the charging / discharging cycles, by longer storage of the gas measurement device in the charging tray and the lazy battery effect.
 CC_{CH_4} =with energy-saving operation at measuring range 0-100%LEL CH₄